

Test Number	Result	ID	Test Time	Test Date	Check Result	Check Time	Check Date	Unit Serial Number
297	0.158		10:14 PM	04/14/2022	0.09	8:34 AM	04/08/2022	16040119
298	0.159		10:17 PM	04/14/2022	0.09	8:34 AM	04/08/2022	16040119
342	0.278		11:35 AM	04/16/2022	0.106	8:35 AM	04/08/2022	16040140
343	0.249		11:38 AM	04/16/2022	0.106	8:35 AM	04/08/2022	16040140
591	0.098		3:12 AM	04/10/2022	0.1	8:20 AM	04/08/2022	16040162
592	0.088		3:14 AM	04/10/2022	0.1	8:20 AM	04/08/2022	16040162
593	0.193		11:53 PM	04/10/2022	0.1	8:20 AM	04/08/2022	16040162
594	0.202		11:56 PM	04/10/2022	0.1	8:20 AM	04/08/2022	16040162
595	0.117		10:35 PM	04/13/2022	0.1	8:20 AM	04/08/2022	16040162
596	0.118		10:38 PM	04/13/2022	0.1	8:20 AM	04/08/2022	16040162
597	0.178		12:04 AM	04/16/2022	0.1	8:20 AM	04/08/2022	16040162
598	0.173		12:06 AM	04/16/2022	0.1	8:20 AM	04/08/2022	16040162
599	0.078		2:59 AM	04/18/2022	0.1	8:20 AM	04/08/2022	16040162
600	0.083		3:01 AM	04/18/2022	0.1	8:20 AM	04/08/2022	16040162
510	0		2:09 AM	04/08/2022	0.101	1:16 PM	03/31/2022	16040163
511	0		1:51 AM	04/18/2022	0.094	8:57 AM	04/08/2022	16040163
335	0.205		6:50 PM	04/12/2022	0.095	9:03 AM	04/08/2022	16080013
336	0.202		6:52 PM	04/12/2022	0.095	9:03 AM	04/08/2022	16080013
204	0.179		9:59 PM	04/15/2022	0.101	9:11 AM	04/08/2022	16080015
205	0.173		10:01 PM	04/15/2022	0.101	9:11 AM	04/08/2022	16080015
206	0.152		1:49 AM	04/16/2022	0.101	9:11 AM	04/08/2022	16080015
207	0.155		1:52 AM	04/16/2022	0.101	9:11 AM	04/08/2022	16080015
293	0.115		6:35 AM	04/09/2022	0.102	9:13 AM	04/08/2022	16080027
294	0.118		6:37 AM	04/09/2022	0.102	9:13 AM	04/08/2022	16080027
224	0.075		10:38 PM	04/15/2022	0.096	9:45 AM	04/08/2022	16080064
225	0.071		10:41 PM	04/15/2022	0.096	9:45 AM	04/08/2022	16080064
226	0		5:38 PM	04/17/2022	0.096	9:45 AM	04/08/2022	16080064
361	0		11:40 PM	04/09/2022	0.1	9:51 AM	04/08/2022	16080069
362	0.122		3:05 AM	04/10/2022	0.1	9:51 AM	04/08/2022	16080069
363	0.132		3:08 AM	04/10/2022	0.1	9:51 AM	04/08/2022	16080069
176	0.154		10:51 PM	04/16/2022	0.092	10:07 AM	04/08/2022	16080075
177	0.152		10:54 PM	04/16/2022	0.092	10:07 AM	04/08/2022	16080075
178	0.155		1:57 AM	04/18/2022	0.092	10:07 AM	04/08/2022	16080075

CHP SLA PAS-Usage Logs 04-18-2022

179	0.153	1:58 AM	04/18/2022	0.092	10:07 AM	04/08/2022	16080075
180	0	2:10 AM	04/18/2022	0.092	10:07 AM	04/08/2022	16080075
274	0.165	3:04 AM	04/10/2022	0.093	11:29 AM	04/08/2022	16120188
275	0.163	3:06 AM	04/10/2022	0.093	11:29 AM	04/08/2022	16120188
276	0.149	2:01 AM	04/15/2022	0.093	11:29 AM	04/08/2022	16120188
277	0.157	2:04 AM	04/15/2022	0.093	11:29 AM	04/08/2022	16120188
278	0.073	10:45 PM	04/15/2022	0.093	11:29 AM	04/08/2022	16120188
279	0.071	10:47 PM	04/15/2022	0.093	11:29 AM	04/08/2022	16120188
366	0	1:47 AM	04/13/2022	0.094	11:39 AM	04/08/2022	16120194
367	0.027	1:48 AM	04/17/2022	0.094	11:39 AM	04/08/2022	16120194
368	0.027	1:50 AM	04/17/2022	0.094	11:39 AM	04/08/2022	16120194
612	0.487	3:55 PM	04/10/2022	0.104	11:45 AM	04/08/2022	16120195
613	0.472	3:59 PM	04/10/2022	0.104	11:45 AM	04/08/2022	16120195
614	0.17	7:22 AM	04/17/2022	0.104	11:45 AM	04/08/2022	16120195
615	0.161	7:24 AM	04/17/2022	0.104	11:45 AM	04/08/2022	16120195
616	0.136	11:24 PM	04/17/2022	0.104	11:45 AM	04/08/2022	16120195
617	0.128	11:27 PM	04/17/2022	0.104	11:45 AM	04/08/2022	16120195
422	0	3:08 PM	04/09/2022	0.093	11:51 AM	04/08/2022	16120197
423	0.119	3:39 AM	04/10/2022	0.093	11:51 AM	04/08/2022	16120197
424	0.124	3:41 AM	04/10/2022	0.093	11:51 AM	04/08/2022	16120197
425	0.076	2:42 AM	04/16/2022	0.093	11:51 AM	04/08/2022	16120197
426	0.137	3:06 AM	04/16/2022	0.093	11:51 AM	04/08/2022	16120197
427	0.13	3:10 AM	04/16/2022	0.093	11:51 AM	04/08/2022	16120197
92	0.168	1:04 AM	04/17/2022	0.094	12:04 PM	04/08/2022	17170066
93	0.165	1:06 AM	04/17/2022	0.094	12:04 PM	04/08/2022	17170066

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 16080071
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
04/18/2022	1149	34.0°	21BP41143	.100	22.3		103				E. Mc Elroy	16193	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.
 Destroy Previous Editions

