

CHP SLA PAS Usage Logs 05-16-2022

Test Number	Result	ID	Test Time	Test Date	Check Result	Check Time	Check Date	Unit Serial Number
301	0.096		4:28 AM	05/06/2022	0.093	9:27 AM	04/28/2022	16040119
302	0.094		4:30 AM	05/06/2022	0.093	9:27 AM	04/28/2022	16040119
303	0.055		12:40 AM	05/16/2022	0.093	10:29 AM	05/06/2022	16040119
304	0.054		12:42 AM	05/16/2022	0.093	10:29 AM	05/06/2022	16040119
354	0		1:49 AM	05/15/2022	0.101	10:29 AM	05/06/2022	16040140
262	0		12:29 AM	05/11/2022	0.101	10:48 AM	05/06/2022	16040141
263	0.24		2:20 AM	05/11/2022	0.101	10:48 AM	05/06/2022	16040141
264	0.237		2:23 AM	05/11/2022	0.101	10:48 AM	05/06/2022	16040141
265	0.105		12:03 AM	05/15/2022	0.101	10:48 AM	05/06/2022	16040141
604	0.17		6:11 PM	05/15/2022	0.104	10:23 AM	05/06/2022	16040162
605	0.147		6:12 PM	05/15/2022	0.104	10:23 AM	05/06/2022	16040162
606	0.162		6:13 PM	05/15/2022	0.104	10:23 AM	05/06/2022	16040162
520	0.177		4:30 AM	05/09/2022	0.105	10:59 AM	05/06/2022	16040163
521	0.169		4:33 AM	05/09/2022	0.105	10:59 AM	05/06/2022	16040163
295	0		3:28 AM	05/08/2022	0.099	11:11 AM	05/06/2022	16080027
296	0.138		10:01 PM	05/11/2022	0.099	11:11 AM	05/06/2022	16080027
364	0.217		12:31 AM	05/10/2022	0.091	12:25 PM	05/06/2022	16080069
365	0.212		12:34 AM	05/10/2022	0.091	12:25 PM	05/06/2022	16080069
655	0.139		3:50 AM	05/07/2022	0.095	2:01 PM	04/28/2022	16080071
656	0.147		3:54 AM	05/07/2022	0.095	2:01 PM	04/28/2022	16080071
657	0.147		2:07 AM	05/14/2022	0.095	2:01 PM	04/28/2022	16080071
658	0.135		2:10 AM	05/14/2022	0.095	2:01 PM	04/28/2022	16080071
659	0		2:12 AM	05/14/2022	0.095	2:01 PM	04/28/2022	16080071
660	0.158		9:59 AM	05/15/2022	0.095	2:01 PM	04/28/2022	16080071
661	0.153		10:01 AM	05/15/2022	0.095	2:01 PM	04/28/2022	16080071
126	0.171		11:34 PM	05/07/2022	0.098	2:49 PM	05/06/2022	16080072
127	0.161		11:37 PM	05/07/2022	0.098	2:49 PM	05/06/2022	16080072
128	0.143		1:25 AM	05/08/2022	0.098	2:49 PM	05/06/2022	16080072
129	0.171		1:08 AM	05/09/2022	0.098	2:49 PM	05/06/2022	16080072
130	0.166		1:10 AM	05/09/2022	0.098	2:49 PM	05/06/2022	16080072
131	0.046		11:51 PM	05/14/2022	0.098	2:49 PM	05/06/2022	16080072
132	0.025		1:09 AM	05/15/2022	0.098	2:49 PM	05/06/2022	16080072
373	0.154		12:30 AM	05/06/2022	0.108	11:07 AM	04/28/2022	16120177

CHP SLA PAS Usage Logs 05-16-2022

374	0.163	12:32 AM	05/06/2022	0.108	11:07 AM	04/28/2022	16120177
375	NEG	8:36 PM	05/07/2022	0.104	3:25 PM	05/06/2022	16120177
376	0.148	8:38 PM	05/07/2022	0.104	3:25 PM	05/06/2022	16120177
377	0.152	8:40 PM	05/07/2022	0.104	3:25 PM	05/06/2022	16120177
378	0	8:23 PM	05/08/2022	0.104	3:25 PM	05/06/2022	16120177
379	0	9:20 PM	05/08/2022	0.104	3:25 PM	05/06/2022	16120177
380	0.136	5:38 PM	05/12/2022	0.104	3:25 PM	05/06/2022	16120177
381	0.137	5:40 PM	05/12/2022	0.104	3:25 PM	05/06/2022	16120177
382	0	7:47 PM	05/15/2022	0.104	3:25 PM	05/06/2022	16120177
287	0	3:38 AM	05/06/2022	0.097	2:04 PM	04/28/2022	16120188
288	0.174	4:00 AM	05/07/2022	0.092	3:27 PM	05/06/2022	16120188
289	0.175	4:02 AM	05/07/2022	0.092	3:27 PM	05/06/2022	16120188
290	0.216	4:53 AM	05/07/2022	0.092	3:27 PM	05/06/2022	16120188
291	0.215	4:55 AM	05/07/2022	0.092	3:27 PM	05/06/2022	16120188
292	0.125	3:34 AM	05/08/2022	0.092	3:27 PM	05/06/2022	16120188
293	0.123	3:36 AM	05/08/2022	0.092	3:27 PM	05/06/2022	16120188
294	0.192	3:20 AM	05/14/2022	0.092	3:27 PM	05/06/2022	16120188
295	0.191	3:22 AM	05/14/2022	0.092	3:27 PM	05/06/2022	16120188
369	0.018	3:49 PM	05/11/2022	0.101	3:30 PM	05/06/2022	16120194
622	0.198	12:47 PM	05/07/2022	0.108	3:31 PM	05/06/2022	16120195
623	0.192	12:49 PM	05/07/2022	0.108	3:31 PM	05/06/2022	16120195
624	0.117	12:44 AM	05/11/2022	0.108	3:31 PM	05/06/2022	16120195
625	0.121	12:46 AM	05/11/2022	0.108	3:31 PM	05/06/2022	16120195
626	0	2:30 AM	05/13/2022	0.108	3:31 PM	05/06/2022	16120195
627	0.022	2:54 AM	05/13/2022	0.108	3:31 PM	05/06/2022	16120195
432	0.179	7:45 PM	05/12/2022	0.095	3:35 PM	05/06/2022	16120197
433	0.163	7:47 PM	05/12/2022	0.095	3:35 PM	05/06/2022	16120197
96	0.04	11:47 PM	05/14/2022	0.096	3:46 PM	05/06/2022	17170066
97	0.116	1:50 AM	05/15/2022	0.096	3:46 PM	05/06/2022	17170066
98	0.106	1:52 AM	05/15/2022	0.096	3:46 PM	05/06/2022	17170066

AREA NAME
 South Los Angeles

AREA NO.
 530

COORDINATOR NAME
 Edward Mc Elroy

I.D. NUMBER
 16193

PAS MAKE/MODEL
 Lifeloc FC20

PAS SERIAL NO.
 16040162

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
				TEST #1 (Optional)	TEST #2 (Optional)		TEST #1 (Optional)	TEST #2 (Optional)	YES	NO							
05/16/2022	1014	34.0°	21BP41143			.100	22.3		1.17			X		.099	E. Mc Elroy	16193	

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Edward Mc Elroy I.D. NUMBER: 16193 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16040119

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	TEMP #1	TEMP #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
		VALUE	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO					
05/16/2022	10:21	34.0°	21BP41143	.100	23,1	.096		X		E. Mc Elroy	16193		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#3

AREA NAME
South Los Angeles

AREA NO.
530

COORDINATOR NAME
Edward Mc Elroy

I.D. NUMBER
16193

PAS MAKE/MODEL
Lifeloc FC20

PAS SERIAL NO.
16090140

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
05/16/2022	1021	34.0°	21BP41143	.100	22.6		105						
			Solution Prepared By:										
			Kyle Mathena										
			Date: 02-03-2021										
			Expiration Date:										
			02-03-2023										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#4

AREA NAME: South Los Angeles

AREA NO.: 530
 COORDINATOR NAME: Edward Mc Elroy

I.D. NUMBER: 16193

PAS MAKE/MODEL: Lifeloc FC20

PAS SERIAL NO.: 16040141

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
05/16/2022	10:25	34.0°	21BP41143	.100	21.8		1.090				E. Mc Elroy	16193	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g. with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions

CHP 202K (Rev. 8-08) OPI 066

AREA NAME: South Los Angeles | AREA NO.: 530 | COORDINATOR NAME: Edward Mc Elroy | I.D. NUMBER: 16193 | PAS MAKE/MODEL: Lifeloc FC20 | PAS SERIAL NO.: 16040161

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
05/16/2022	1032	34.0°	21BP41143	.100	21.6	.100			X		E. Mc Elroy	16193	
Solution Prepared By: Kyle Mathena													
Date: 02-03-2021													
Expiration Date: 02-03-2023													

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions

Chp202K_1221.pdf

AREA NAME	AREA NO.	COORDINATOR NAME	ID. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Edward Mc Elroy	16193	Lifoloc FC20	16040163

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
05/16/2022	1035	34.0°	21BP41143	.100	23,1	.101			X		E. Mc Elroy	16193	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

17

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Edward Mc Elroy I.D. NUMBER: 16193 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080013

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
					TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
05/16/2022	1043	34.0°	21BP41143	.100	22.3		1.096			X		E. Mc Elroy	16193	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 16080014
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO					
05/16/2022	1044	34.0°	21BP41143	.100	23.1		1.096								

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g. with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Edward Mc Elroy
 I.D. NUMBER: 16193
 PAS MAKE/MODEL: Lifeloc FC20
 PAS SERIAL NO.: 16080015

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
05/16/2022	1045	34.0°	21BP41143	.100	22.8	1.00					E. Mc Elroy	16193	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#10

AREA NAME: South Los Angeles | AREA NO.: 530 | COORDINATOR NAME: Edward Mc Elroy | I.D. NUMBER: 16193 | PAS MAKE/MODEL: Lifeloc FC20 | PAS SERIAL NO.: 16080627

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
05/16/2022	1047	34.0°	21BP41143	.100	22.8	.098				X	E. Mc Elroy	16193	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

#11

CHP 202K (Rev. 8-08) OPI 066

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 16080028
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
		TEMP (°C)	LOT NO.		TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO					
05/16/2022	1055	34.0°	21BP41143	.100	21.6		1097				X		E. Mc Elroy	16193	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Edward Mc Elroy I.D. NUMBER: 16193 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080072

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.			TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
05/16/2022	1131	34.0°	21BP41143		.100	21.8		.081		X		.109	E. Mc Elroy	16193	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#18

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Edward Mc Elroy I.D. NUMBER: 16193 PAS MAKE/MODEL: Lifelog FC20 PAS SERIAL NO.: 16080075

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
05/16/2022	1134	34.0°	21BP41143	.100	21.8	.098					E. Mc Elroy	16193	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions

Chp202K_1221.pdf

AREA NAME: South Los Angeles
AREA NO.: 530
COORDINATOR NAME: Edward Mc Elroy
ID. NUMBER: 16193
PAS MAKE/MODEL: Lifelog FC20
PAS SERIAL NO.: 16080077

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
05/16/2022	1136	34.0°	21BP41143	.100	22.1	.102				X	E. Mc Elroy	16193	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 16120195
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
		TEMP (°C)	LOT NO.		TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO					
05/16/2022	1153	34.0°	21BP41143	.100	21.8		1.105				X		E. Mc Elroy	16193	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions

